

Certificate of Compliance

Certificate Number 20070926 - E131021
Report Reference E131021, 2003 August 28
Issue Date 2007 September 26

Page 1 of 1



**Underwriters
Laboratories Inc.**

Issued to: **BW Technologies By Honeywell**

2840 2nd Ave SE
Calgary, AB T2A 7X9 Canada

*This is to certify that
representative samples of:*

Gas and Vapor Detection Equipment

Two-year GasAlertClip Extreme Series, Models GA24XT-H, GA24XT-M, GA24XT-S and GA24XT-X gas monitors, and Three-year GasAlertClip Extreme Series, Models GA36XT-H, GA36XT-M, GA36XT-S and GA36XT-X gas monitors, intrinsically safe for use in Class I, Groups A, B, C and D Hazardous Locations when used with one self-contained 3.6 V lithium battery.

Have been investigated by Underwriters Laboratories Inc.® (UL) or any authorized licensee of UL in accordance with the Standard(s) indicated on this Certificate.


Standard(s) for Safety:


UL 913 Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations, Seventh Edition.
Canadian National Standard C22.2 No. 157-92.

Additional Information:

Marking shall include manufacturer's name or trademark, model number, hazardous location class and group, temperature code, ambient temperature range, the words "Only as to Intrinsic Safety", and the warning "Understand Manual before Operating".

Only those products bearing the UL Classification Mark for the U.S. and Canada should be considered as being covered by UL's Classification and Follow-Up Service and meeting the appropriate U.S. and Canadian requirements.

The UL Classification Mark includes: the UL in a circle symbol  with the word "CLASSIFIED" (as shown); a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) as indicated in the appropriate UL Directory. The

UL Classification Mark for Canada includes: the UL Classification Mark for Canada  with the word "CLASSIFIED" (as shown); a control number (may be alphanumeric) assigned by UL; a statement to indicate the extent of UL's evaluation of the product; and the product category name (product identity) in English, French, or English/French as indicated in the appropriate UL Directory.

Look for the UL Classification Mark on the product

Issued by:

Casey Martin / Associate Project Engineer
Underwriters Laboratories Inc.

Reviewed by:

Joseph Getner / Section Manager
Underwriters Laboratories Inc.