

| | | | | | |
|--------------------------------------|--|----------------|---------------|-----|-----|
| Honeywell Application Note | Usage and Applications of PID Sensors | Date | 08/22/11 | Rev | B11 |
| | | N ^o | 13-392 | | |
| | | By | Jeffrey Emond | | |

Introduction to PID Sensors and VOCs

The photoionization detection (PID) sensor used in the PHD6 multi-gas detector is a broadband sensor primarily designed to detect volatile organic compounds (VOCs), which are carbon containing chemicals that can exist in gaseous form at ambient temperatures. The sensor will also detect certain other toxic gases.

The body of the sensor is made up of a UV (ultraviolet light) lamp and an electrode stack. Internal electronic circuitry, (an ASIC) is used regulate operation of both the lamp and stack. It is also used to compensate the output signal for environmental conditions. There is also an onboard smart chip used to store information such as serial number, temperature compensation, calibration dates, etc.

When the target gas enters the sensor's sensing chamber, the photons emitted from the UV lamp break up the molecule into two ions, one positively charged, one negatively charged. The electric field created by the anode and cathode in the stack attracts the ions, which results in an electric current proportional to the concentration of the gas. This current is then measured by the circuitry in the sensor, and a signal is sent to the instrument, which then reports and records the concentration.

The PHD6 PID sensor is designed for use in either diffusion or pumped sampling modes. The advantage of a pump is generally quicker response and recovery. While for diffusive sampling, retention of higher ozone concentrations, generated in the ionization chamber, can act as cleaning agent for VOC residues on electrode & lamp window surfaces.

Linear and Working Ranges

PID sensors are broadband in nature. This means that they are inherently non-specific. Any gas or vapor that is ionized at the UV lamp energy level will give a response.

It must be understood that the selection of a particular VOC or gas from the onboard PID library in the PHD6 does **not** imply that the detector will only respond to that material. It only means that the sensitivity scale (and default alarms) has been set to approximate the target material. Regardless of the library material selected, the PID sensor always remains broadband in nature and therefore will respond to any gases/vapors in the ambient environment that are present and are ionized at the UV lamp energy. This consideration is particularly important when trace or hard to detect materials (higher correction factor (CF)) are present in highly contaminated backgrounds. In this case the PID would be a poor choice for detection of the target gas/vapor.

The PID is traditionally calibrated using 100 PPM isobutylene (IB) because its sensitivity is near the midpoint of most VOCs, and it is non-toxic and non-flammable at the low concentrations used for calibration. Being the primary reference, its correction factor (CF) is defined as 1.0. CF's for all photoionizable gases and vapors are calculated based on their sensitivity as compared to isobutylene.

The PID sensor has a linear response up to about 2000 PPM for materials with an ionization potential similar to isobutylene (CF~1). Sensitivity decreases when concentrations exceed this value since some of the gas molecules may become shielded from the UV light and fail to ionize. If the instrument will be used in this type of envi-

ronment, it must be calibrated with a high concentration calibration gas. As a result, readings at the low end may become less accurate.

The PID channel will allow a reading of up to 3000 PPM, since for materials that are less ionizable than isobutylene (CF>1), the linear range may be extended up to this amount. Please consult with Honeywell application support at (800) 711-6776 prior to using the PHD6 to monitor extremely high concentrations of any gas/vapor.

The PHD6 PID channel may be configured by skilled operators in a way to maximize the working range. To highlight this case consider as a specific example, ammonia (CF = 8.5). If the PID channel is left on an isobutylene setting/scale, the detector subsequently used in an ammonia containing atmosphere could read up to 3,000 PPM. On this scale, therefore, 3,000 PPM IB scale X 8.5 CF = 25,500 PPM ammonia. Theoretically it is possible to read up to 25,500 PPM ammonia. However, as stated previously, the linear range of the PID sensor is normally 2000 PPM (isobutylene) which would place the linear maximum for ammonia more on the order of 17,000 PPM. Alternately if the PID channel is directly configured for an ammonia scale, then the channel range will be limited to 0-3000 PPM ammonia. Keep in mind about the former case that the end user and not the detector would have to apply the correction factor to the displayed value to get the actual level of ammonia. The alarm setpoints would also have to be adjusted as they would be relevant to isobutylene, not ammonia. Extreme caution must be practiced when using the PHD6 in this mode to detect VOC's or other gases with a CF much different than 1. Failure to do so can cause over-exposure, which can re-

| | | | | | |
|--------------------------------------|--|----------------|---------------|-----|-----|
| Honeywell Application Note | Usage and Applications of PID Sensors | Date | 08/22/11 | Rev | B11 |
| | | N ^o | 13-392 | | |
| | | By | Jeffrey Emond | | |

sult in serious personal injury or death.

Resolution / Clamping

The best resolution you should expect from the PID sensor in the PHD6 is the sensor supplier's specified resolution of 0.1 PPM, for materials with a CF close to 1, when the sensor and detector are new and freshly calibrated and operating in benign environmental conditions. A safer rule-of-thumb is to expect about 1 PPM of resolution on the IB scale under moderate conditions and not to depend on alarm setpoints less than about 5 times that, or 5 PPM. If a CF different than 1 is used, scale these values appropriately as described above. Lower alarm settings can be used by skilled safety professionals taking appropriate precautions such as calibrating directly with the VOC/gas of interest. Failure to do so can cause over-exposure, which can result in serious personal injury or death.

In addition, the PHD6 utilizes a clamping mechanism. The standard clamp is 1 PPM for the VOC channel. This means that readings between -0.9 and +0.9 PPM are displayed as 0 PPM. This is not a problem with a gas such as isobutylene, whose ceiling alarm is 1800 PPM. But it can be an issue if the alarm is less than 1 PPM (again, not recommended). In this case, the PHD6 will automatically adjust the clamp to equal the ceiling alarm level. So if the ceiling alarm level for a particular target gas is 0.5 PPM, readings between -0.4 and +0.4 PPM will be displayed as 0 PPM. The PID is active and detecting gas when clamped, only the reading is suppressed.

Methane Inhibition

CAUTION: Some gases such as Methane and ethane can interfere with PID response and cause it to read low.

The PID sensor response to target gases/vapors is measured in laboratory air, with 20.9% oxygen, balance nitrogen. Some gases absorb UV light without causing any PID response (e.g., methane, ethane). In ambient atmospheres where these gases are present, the measured concentration of target gas will be less than is actually present. Methane absorbs UV strongly, so for accurate measurements in methane containing atmospheres, calibrate with a gas containing the expected methane concentration. 50% LEL methane (2.5% vol.) can reduce the reading by up to 50%.

Gases such as nitrogen, oxygen, argon and helium do not absorb UV and do not affect PID sensor response.

Cross-Sensitivity with Duo-Tox Sensors

The CO and H₂S channels of the Duo-Tox sensor used in the PHD6 may exhibit cross-sensitivity to VOCs. When exposed to 100 PPM isobutylene, the CO reading may rise as high as 60 PPM. The H₂S channel, when exposed to certain VOC's such as alcohols and other chemically active vapors may exhibit a negative cross-interference.

Dedicated CO gas sensors have built-in filters which help prevent this type of cross-sensitivity. Over time, these filters degrade due to age and exposure. In addition, the effectiveness of the filters depends largely on the VOC being detected. Single CO sensors have better filtration for VOCs than Duo-Tox sensors. In fact, a new CO sensor usually will have essentially no cross-interference with 100 PPM isobutylene. Dedicated H₂S sensors have another intrinsic means (not an internal filter) to resist VOC cross-interference. Again, this is not the case for the DuoTox sensor design.

Therefore when an instrument configuration of PID, CO and H₂S is necessary, for the best performance, Honeywell strongly recommends using separate CO and H₂S sensors (p/ns 54-54-01 and 54-54-02) in place of the DuoTox sensor (p/n 54-54-14).

Questions regarding specific VOC cross-sensitivity on any toxic gas sensor configurations should be referred to Honeywell application support at (800) 711-6776.

Correction or Response Factors

PID sensors are broadband sensors. They will detect many different compounds, and few of these compounds will show the same response when detected by the sensor.

The calibration standard for PID sensors is 100 PPM isobutylene. Although the most accurate way to detect VOCs is by calibrating directly to the target gas, most common VOCs have a approximate correction factor (CF) also referred to as a "response factor" (RF) published by the sensor manufacturer that can be applied to the sensor reading so that an approximate value can be determined when calibrated to isobutylene.

Correction factors in the PHD6 on-board PID library for various, common VOCs and gases should be considered as approximate. The PHD6 with PID has been fully tested and validated only for performance with isobutylene. For other materials requiring verified accuracy it is necessary to calibrate the detector to the gas/vapor to be monitored directly. Further, if using directed flow (remote sample draw) and/or physical conditions in the field that differ from ambient, to perform calibrations as close to the physical and actual setup conditions as possible.

| | | | | | |
|--------------------------------------|--|----------------|---------------|-----|-----|
| Honeywell Application Note | Usage and Applications of PID Sensors | Date | 08/22/11 | Rev | B11 |
| | | N ^o | 13-392 | | |
| | | By | Jeffrey Emond | | |

Initially, two factors are very important for consideration when using a PID to monitor for a particular VOC/gas. These are the ability of the PID to detect as noted by the CF and the toxicity as indicated by the permissible exposure limits (PEL's) or published alarm levels. PID sensitivity is rated or normalized in reference to isobutylene, the calibration gas standard, where CF=1. A higher value indicates that the PID has a more difficult time in ionizing/sensing the material. PEL's or alarm levels that are low, generally less than 10 PPM, indicate materials that are very toxic. In general the higher the CF and the lower the PEL's the more care that should be taken when monitoring these materials and protecting operators. General guidelines are that if the CF \geq PEL (lowest PEL value), the detector must not be used as the primary/only safety equipment for personnel, and further, steps should be taken to protect those affected with additional PPE, as applicable to the specific material.

Further, when a compound is selected from the library, the PHD6 PID channel will automatically show a 7 character abbreviation for the scale which has been chosen. Alarms will also be automatically set when a choice is made from the library.

Note that these "default" alarms, automatically set, as shown in the attached table were selected as generally the most conservative levels based on a variety of published standards. However, local laws and regulations must be taken into account, and the user must set them accordingly. Again, an alarm set-point of less than the equivalent of 5 PPM (IB scale) is not recommended. For example, with a CF=2, alarms less than $2 \times 5 = 10$ PPM are not recommended.

Sampling

The PHD6 with PID is set-up to be used in either diffusion or via sample draw methods, with a pump. It is important to use only the tubing or other remote sampling components supplied by Honeywell. It is particularly important to avoid the use of common vinyl (PVC) tubing in sampling systems as this material is known to strongly absorb many classes of VOC's resulting in low/no response and/or long response and recovery times. Further, the more chemically reactive or toxic materials, if they must be sampled remotely, may require the use of FEP-lined sampling lines. Contact Honeywell, technical support (800-711-6776) for details associated with these applications.

Most VOC's and many gases are much heavier than air. Therefore proper sampling becomes a critical matter in order to get proper readings. In stagnant space conditions gas/vapor concentration can be very stratified. It is therefore particularly important to monitor throughout the space from highest to lowest points. Note further consideration should be made to the case that when a space is disturbed, the concentration profile of the gas/vapor may suddenly shift to a significant degree.

Many organic compounds are considered "heavy" or have a high boiling point. In particular, and most common, are heavier petroleum distillates and fuels. These materials can be easily absorbed onto sampling surfaces and therefore may give low and sluggish responses. In these cases it is best to use the shortest sampling lines possible and to allow for longer sampling times to give a proper response. Further given their higher boiling point, their vapors are more apt to condense on sampling lines and components if these are cooler than the sampled environment. In this case it is always best to ensure that the

detector and sampling lines are kept warmer than the sampled environment. Examples of such materials include, but are not limited to:

Biphenyl
Diphenylene oxide
Diesel Fuels.
Diethyl benzene
Ethanolamine
Ethylene glycol
Naphtha(s) – heavier grades, including hydro-treated
Propylene carbonate

Certain VOC's and phosphine are known to interact with UV in such a way as to break down into products that are known or more apt to coat the UV lamp. If monitoring for these, or if they are present, will require more frequent cleaning of the PID lamp will be required. These materials include:

Phosphine (echem sensor preferred)
Volatile Silicones
Organic acids, (acetic, propionic)

Complex VOC Gas Mixtures

A PID cannot distinguish between different gases in a mixture, and its sensitivity to each gas differs. The displayed reading represents the total concentration of all photo-ionizable gases present in the sample.

For a VOC mix of known composition, if the total concentration is within the linear range the PID, then it is reasonable to assume that the concentrations are additive without interference between the different VOCs. This additive result is based on each compound's correction factor and can be approximated by:

$$CF(\text{mix}) = 1 / [(a/CF(A) + b/CF(B) + c/CF(C) \dots)]$$

where CF (mix) is the correction factor for a gas mix containing PID

| | | | | | |
|--------------------------------------|--|------|---------------|-----|-----|
| Honeywell Application Note | Usage and Applications of PID Sensors | Date | 08/22/11 | Rev | B11 |
| | | Nº | 13-392 | | |
| | | By | Jeffrey Emond | | |

detectable gases A, B, C... , in relative proportions a: b: c...

As an example a paint manufacturer formulates with a mixed solvent containing 25% methyl ethyl ketone, 10% toluene and 65% isopropanol. The questions then become:

1) What is the CF for this mix?

From the formula above, $CF(mix) = 1 / [(0.25/0.77) + (0.10/0.55) + (0.65/4.35)]$
Therefore $CF(mix) = 1.52$

2) What are the alarm setpoint(s)?

Using a very similar formula to the above:

$$PEL(mix) = 1 / [(a/PEL(A) + b/PEL(B) + c/PEL(C)...]$$

$$Ceiling/Danger\ PEL = 1 / [(0.25/200) + (0.10/300) + (0.65/200)]$$

Ceiling/Danger PEL = 207 PPM

Similarly PEL(mix) for STEL and TWA would become respectively, 320 PPM and 154 PPM.

3) How can this information be programmed into the PhD6 PID channel?

Through the PhD6 menu select "custom" from the compound library. Into the respective data fields configure the CF value as 1.52,

alarm setpoints for Ceiling/Danger as 207 PPM, STEL as 320 PPM and TWA as 154 PPM. In this way the detector can continue to be calibrated with isobutylene and the custom scale will correspond to this solvent mix.

There can be a case where the volatile materials in a mixture are known but their exact proportions are not. For this application we can use the above three materials and simply assume composition is unknown. We can then construct a table to show how the PID would respond at a PEL value as follows:

| Material | CF | PEL (TWA) Material | PEL (TWA) Isobutylene Scale |
|---------------------|------|-----------------------|--------------------------------|
| Methyl Ethyl Ketone | 0.77 | 200 | 260 |
| Toluene | 0.51 | 50 | 98 |
| Isopropanol | 4.35 | 200 | 46 |

It can be seen from the above table that even though isopropanol is only moderately toxic, it is detected least well by the PID sensor operating

at 10.6 eV (highest CF). Therefore if the TWA alarm setpoint is changed to 46 PPM, then all other vapors should be well below their

TWA-PELs. A similar set of calculations can be done for all other PEL types (ceiling & STEL).

| | | | | | |
|--------------------------------------|--|------|---------------|-----|-----|
| Honeywell Application Note | Usage and Applications of PID Sensors | Date | 08/22/11 | Rev | B11 |
| | | Nº | 13-392 | | |
| | | By | Jeffrey Emond | | |

PhD6 PID Sensor Compound Library with Correction Factors and Default Alarms

| Compound | | | | | Alarms - PPM | | |
|-------------------------|----------------|---|---------------------|--------------------------|----------------------------|-------------|------------|
| Proper Name | CAS No. | Formula | 7-Char Abbr. | Correction Factor | Danger/ Ceiling | STEL | TWA |
| Custom | | | VOC | 1 | 0 | 0 | 0 |
| Acetaldehyde | 75-07-0 | C ₂ H ₄ O | Acetald | 4.86 | 25 | 150 | 100 |
| Acetone | 67-64-1 | C ₃ H ₆ O | Acetone | 0.72 | 500 | 750 | 500 |
| Acetophenone | 98-86-2 | C ₈ H ₈ O | Acetoph | 0.59 | 10 | 30 | 10 |
| Ammonia | 7664-41-7 | NH ₃ | NH3-PID | 8.5 | 25 | 35 | 25 |
| Amyl Acetate | 628-63-7 | C ₇ H ₁₄ O ₂ | AmylAce | 1.8 | 50 | 100 | 50 |
| Amyl Alcohol | 75-85-4 | C ₅ H ₁₂ O | AmylAlc | 3.2 | 1200 | 0 | 0 |
| Aniline | 62-53-3 | C ₇ H ₇ N | Aniline | 0.5 | 2 | 6 | 2 |
| Anisole | 100-66-3 | C ₇ H ₈ O | Anisole | 0.47 | 0 | 0 | 0 |
| Benzaldehyde | 100-52-7 | C ₇ H ₆ O | BenzAld | 0.86 | 2 | 4 | 2 |
| Benzyl Alcohol | 100-51-6 | C ₇ H ₈ O | BenzAlc | 1.25 | 10 | 0 | 10 |
| Bromobenzene | 108-86-1 | C ₆ H ₅ Br | BromBnz | 0.7 | 500 | 0 | 0 |
| Butoxyethanol,2- | 75-25-2 | CHBr ₃ | MeBr ₃ | 2.8 | 20 | 60 | 20 |
| Butanol | 71-36-3 | C ₄ H ₁₀ O | Butanol | 4.01 | 50 | 60 | 20 |
| Butyl Acetate | 123-86-4 | C ₆ H ₁₂ O ₂ | ButAcet | 2.42 | 150 | 200 | 150 |
| Butyl Acrylate | 141-32-2 | C ₇ H ₁₂ O ₂ | ButAcrl | 1.5 | 2 | 6 | 2 |
| Carbon Disulfide | 75-15-0 | CS ₂ | CS ₂ | 1.4 | 4 | 12 | 4 |
| Chlorotrifluoroethylene | 79-38-9 | C ₂ ClF ₃ | C2ClF3 | 1 | 5 | 20 | 5 |
| Chlorobenzene | 108-90-7 | C ₆ H ₅ Cl | ClBenzn | 0.45 | 10 | 30 | 10 |
| Cresols | | C ₇ H ₈ O | Cresols | 1.05 | 5 | 15 | 5 |
| Cumene | 98-82-8 | C ₉ H ₁₂ | Cumene | 0.59 | 50 | 150 | 50 |
| Cyclohexane | 110-82-7 | C ₆ H ₁₂ | Cyhexan | 1.16 | 100 | 300 | 100 |
| Cyclohexanol | 108-93-0 | C ₆ H ₁₂ O | CyHexol | 2.91 | 50 | 150 | 50 |
| Cyclohexanone | 108-94-1 | C ₆ H ₁₀ O | Cyhexon | 1.04 | 25 | 50 | 25 |
| Cyclopentane | 287-92-3 | C ₅ H ₁₀ | Cypentn | 4 | 600 | 0 | 600 |
| Decane | 124-18-5 | C ₁₀ H ₂₂ | Decane | 1.04 | 800 | 0 | 0 |
| Diacetone Alcohol | 123-42-2 | C ₆ H ₁₂ O ₂ | DiacAlc | 0.8 | 50 | 150 | 50 |
| Dibromoethane,1,2- | 106-93-4 | C ₂ H ₄ Br ₂ | DiBrEt | 2 | 30 | 15 | 20 |
| Dichlorobenzene,1,2- | 95-50-1 | C ₆ H ₄ Cl ₂ | DiClBen | 0.5 | 25 | 50 | 25 |
| Dichloroethylene,1,1- | 75-35-4 | C ₂ H ₂ Cl ₂ | DiClEt | 0.95 | 5 | 15 | 5 |
| Dichloroethylene,1,2- | 107-06-2 | C ₂ H ₄ Cl ₂ | DiClEt | 0.7 | 200 | 0 | 200 |
| Diesel Fuel #1 | 68334-30-5 | | Diesel1 | 0.9 | 10 | 0 | 10 |
| Diesel Fuel #2 | 68334-30-5 | | Diesel2 | 0.75 | 10 | 0 | 10 |
| Diethylamine | 109-89-7 | C ₄ H ₁₁ N | DiEtAmn | 1 | 10 | 25 | 10 |
| Dimethoxymethane | 109-87-5 | C ₃ H ₈ O ₂ | DiMeOMe | 1.4 | 1000 | 0 | 1000 |
| Dimethyldisulfide | 624-92-0 | C ₂ H ₆ S ₂ | DiMeDiS | 0.23 | 0.5 | 0 | 0.5 |
| Dimethylacetamide N,N- | 127-19-5 | C ₄ H ₉ NO | DiMeAct | 1.3 | 10 | 30 | 10 |
| Dimethylaniline N,N- | 121-69-7 | C ₈ H ₁₁ N | DiMeAnl | 0.6 | 5 | 10 | 5 |
| Dimethylformamide | 68-12-2 | C ₃ H ₇ NO | DiMeFrm | 0.9 | 10 | 30 | 10 |
| Dioxane, 1,4- | 123-91-1 | C ₄ H ₈ O ₂ | Dioxane | 1.5 | 20 | 60 | 20 |
| Divinylbenzene | 1321-74-0 | C ₁₀ H ₁₀ | VinStyr | 0.4 | 10 | 30 | 10 |

| Compound | | | | | Alarms - PPM | | |
|------------------------|----------------|---|---------------------|--------------------------|----------------------------|-------------|------------|
| Proper Name | CAS No. | Formula | 7-Char Abbr. | Correction Factor | Danger/ Ceiling | STEL | TWA |
| Ethanol | 64-17-5 | C ₂ H ₆ O | Ethanol | 8.72 | 1000 | 0 | 1000 |
| Ethyl Acetate | 141-78-6 | C ₄ H ₈ O ₂ | EtAcet | 3.63 | 400 | 0 | 400 |
| Ethyl Acetoacetate | 141-97-9 | C ₆ H ₁₀ O ₃ | EtAcoAc | 0.9 | 1400 | 0 | 0 |
| Ethyl Acrylate | 140-88-5 | C ₅ H ₈ O ₂ | EtAcryl | 2 | 5 | 15 | 5 |
| Ethyl Ether | 60-29-7 | C ₄ H ₁₀ O | EtEther | 1.2 | 400 | 500 | 400 |
| Ethyl Mercaptan | 75-08-1 | C ₂ H ₆ S | EtSH | 0.7 | 10 | 1.5 | 0.5 |
| Ethylbenzene | 100-41-4 | C ₈ H ₁₀ | EtBenzn | 0.54 | 100 | 125 | 100 |
| Ethylene | 74-85-1 | C ₂ H ₄ | Ethene | 8 | 200 | 600 | 200 |
| Ethylene Glycol | 107-21-1 | C ₂ H ₆ O ₂ | EtGlycl | 20 | 40 | 0 | 0 |
| Furfural | 98-01-1 | C ₅ H ₄ O ₂ | Furfurl | 1.39 | 2 | 6 | 2 |
| Gasoline | 8006-61-9 | | Gasolne | 1.05 | 300 | 500 | 300 |
| Heptane | 142-82-5 | C ₇ H ₁₆ | Heptane | 2.06 | 400 | 500 | 400 |
| Hexane, n- | 110-54-3 | C ₆ H ₁₄ | Hexane | 4.2 | 50 | 0 | 50 |
| Hydrogen Sulfide | 7783-06-4 | H ₂ S | H2S-PID | 4 | 10 | 15 | 10 |
| Indene | 95-13-6 | C ₉ H ₈ | Indene | 0.46 | 10 | 30 | 10 |
| Iodomethane | 74-99-4 | CH ₃ I | IodMetn | 0.4 | 2 | 6 | 2 |
| Isoamyl Acetate | 123-92-2 | C ₇ H ₁₄ O ₂ | IamlAct | 1.6 | 50 | 100 | 50 |
| Isobutanol | 78-83-1 | C ₄ H ₁₀ O | Ibutanl | 3.5 | 50 | 150 | 50 |
| IsobutylAcetate | 110-19-0 | C ₆ H ₁₂ O ₂ | IbtlAct | 2.26 | 150 | 450 | 150 |
| Isobutylene | 115-11-7 | C ₄ H ₈ | Ibutyln | 1 | 1800 | 0 | 0 |
| Isooctane | 540-84-1 | C ₈ H ₁₈ | Ioctane | 1.08 | 1100 | 0 | 0 |
| Isopentane | 78-78-4 | C ₅ H ₁₂ | Ipentan | 6 | 1400 | 0 | 0 |
| Isophorone | 78-59-1 | C ₉ H ₁₄ O | Iphoron | 0.75 | 5 | 0 | 25 |
| Isoprene | 78-79-5 | C ₅ H ₈ | Iprene | 0.7 | 1500 | 0 | 50 |
| Isopropanol | 67-63-0 | C ₃ H ₈ O | Ipropnl | 4.35 | 200 | 400 | 200 |
| Isopropyl Acetate | 108-21-4 | C ₅ H ₁₀ O ₂ | IplAcet | 2.2 | 100 | 200 | 100 |
| Isopropyl Ether | 108-20-3 | C ₆ H ₁₄ O | IplEthr | 0.8 | 250 | 310 | 250 |
| Isopropylamine | 75-31-0 | C ₃ H ₉ N | Iplamin | 0.9 | 5 | 10 | 5 |
| Jet A Fuel | | | JetA | 0.65 | 34 | 0 | 34 |
| Jet B Fuel | | | JetB | 0.75 | 30 | 0 | 30 |
| JP-4 Fuel | | | JP-4 | 0.75 | 30 | 0 | 30 |
| JP-5 Fuel | | | JP-5 | 0.65 | 29 | 0 | 29 |
| Compound | | | | | Alarms - ppm | | |
| Proper Name | CAS No. | -- | Kerosen | 0.83 | 34 | 0 | 34 |
| Mesitylene | 108-67-8 | C ₉ H ₁₂ | Mestyln | 0.34 | 25 | 75 | 25 |
| Mesityl Oxide | 141-79-7 | C ₆ H ₁₀ O | MestyLO | 0.47 | 15 | 25 | 15 |
| Methacrylic Acid | 79-41-4 | C ₄ H ₆ O ₂ | MeAcryA | 2.3 | 20 | 60 | 20 |
| Methoxyethanol 1,2- | 109-86-4 | C ₃ H ₈ O ₂ | MeOetnl | 2.7 | 5 | 15 | 5 |
| Methoxypropanol 1,1- | | | MeOprnl | 3 | 100 | 150 | 100 |
| Methyl Acetate | 79-20-9 | C ₃ H ₆ O ₂ | MeAcet | 5.19 | 200 | 250 | 200 |
| Methylacetoacetate | 105-45-3 | C ₅ H ₈ O ₃ | MeAcoAc | 1.1 | 0 | 0 | 0 |
| Methyl Benzoate | 93-58-3 | C ₈ H ₈ O ₂ | MeBenzo | 0.93 | 0 | 0 | 0 |
| Methyl Benzyl Alcohol | 589-18-4 | C ₈ H ₁₀ O | MeBeAlc | 0.8 | 0 | 0 | 0 |
| Methyl Ethyl Ketone | 78-93-3 | C ₄ H ₈ O | MEK | 0.77 | 200 | 300 | 200 |
| Methyl Isobutyl Ketone | 108-10-1 | C ₆ H ₁₂ O | MIBK | 0.8 | 50 | 75 | 50 |
| Methyl Methacrylate | 80-62-6 | C ₅ H ₈ O ₂ | MMA | 1.6 | 50 | 100 | 50 |
| Methyl t-Butyl Ether | 1634-04-4 | C ₅ H ₁₂ O | MTBE | 0.8 | 50 | 150 | 50 |

| | | | | | | | |
|---------------------------|-------------------------|---|----------------------------|---------------------------------|-------------------------------|--------------------|-------------------|
| Methylamine | 74-89-5 | CH ₅ N | MeAmine | 1.4 | 5 | 15 | 5 |
| <u>Compound</u> | | | | | <u>Alarms - PPM</u> | | |
| <u>Proper Name</u> | <u>CAS No.</u> | <u>Formula</u> | <u>7-Char Abbr.</u> | <u>Correction Factor</u> | <u>Danger/ Ceiling</u> | <u>STEL</u> | <u>TWA</u> |
| Methyl Pyrrolidinone | 872-50-4 | C ₅ H ₉ NO | MePyrrd | 0.9 | 10 | 0 | 10 |
| Mineral Spirits | 8030-30-6 8032-32-4 | -- | MinSprt | 0.8 | 100 | 0 | 100 |
| Naphtha | 8030-30-6 8052-41-3 | -- | Naphtha | 0.8 | 100 | 0 | 100 |
| Naphthalene | 91-20-3 | C ₁₀ H ₈ | Naphthl | 0.44 | 10 | 15 | 10 |
| Nitric Oxide | 10102-43-9 | NO | NO-PID | 8 | 25 | 75 | 25 |
| Nonane, n- | 111-84-2 | C ₉ H ₂₀ | Nonane | 1.27 | 200 | 600 | 200 |
| Octane | 111-65-9 | C ₈ H ₁₈ | Octane | 1.59 | 300 | 375 | 300 |
| Pentane, n- | 109-66-0 | C ₅ H ₁₂ | Pentane | 7.89 | 600 | 750 | 600 |
| Pentyl Alcohol | 71-41-0 | C ₅ H ₁₂ O | PentAlc | 3.2 | 1200 | 0 | 0 |
| 2-Pentanone | 107-87-9 | C ₅ H ₁₀ O | Pentone | 0.79 | 200 | 250 | 200 |
| Petrol | 8006-61-9 | | Petrol | 1.05 | 300 | 500 | 300 |
| Phenol | 108-95-2 | C ₆ H ₆ O | Phenol | 1.2 | 5 | 15 | 5 |
| Picoline,2- | 109-06-8 | C ₆ H ₇ N | Picolne | 0.57 | 2 | 5 | 2 |
| Picoline,3- | 108-99-6 | C ₆ H ₇ N | Picolne | 0.9 | 2 | 5 | 2 |
| Pinene, α- & β- | 2437-95-8 18172-67-3 | C ₁₀ H ₁₆ | Pinene | 0.32 | 20 | 60 | 20 |
| Propanol | 71-23-8 | C ₃ H ₈ O | Propanl | 4.8 | 200 | 250 | 200 |
| Propionaldehyde | 123-38-6 | C ₃ H ₆ O | Propald | 1.68 | 20 | 0 | 20 |
| Propionic Acid | 79-09-4 | C ₃ H ₆ O ₂ | PropAcid | 8 | 10 | 30 | 10 |
| Propyl Acetate | 109-60-4 | C ₅ H ₁₀ O ₂ | PropAct | 2.5 | 200 | 250 | 200 |
| Propylene | 115-07-01 | C ₃ H ₆ | Propene | 1.4 | 0 | 0 | 0 |
| Pyridine | 110-86-1 | C ₅ H ₅ N | Pyridne | 0.75 | 1 | 3 | 1 |
| Styrene | 100-42-5 | C ₈ H ₈ | Styrene | 0.44 | 20 | 40 | 20 |
| t-Butyl Alcohol | 75-65-0 | C ₄ H ₁₀ O | TB Alc | 3.4 | 100 | 150 | 100 |
| t-Butyl Mercaptan | 76-66-1 | C ₄ H ₁₀ S | TriBuSH | 0.55 | 0 | 0 | 0 |
| t-Butylamine | 75-64-9 | C ₄ H ₁₁ N | TBAmine | 0.71 | 0 | 0 | 0 |
| Tetrachloroethylene | 127-18-4 | C ₂ Cl ₄ | TetrClE | 0.7 | 25 | 100 | 25 |
| Tetrahydrofuran | 109-99-9 | C ₄ H ₈ O | THF | 1.55 | 50 | 100 | 50 |
| Thiophene | 110-02-1 | C ₄ H ₄ S | Thiophn | 0.47 | 1500 | 0 | 0 |
| Toluene | 108-88-3 | C ₇ H ₈ | Toluene | 0.51 | 300 | 150 | 50 |
| Trichloroethylene | 79-01-6 | C ₂ HCl ₃ | TriClE | 0.65 | 200 | 100 | 50 |
| Trimethylamine | 75-50-3 | C ₃ H ₉ N | TrMeAmn | 0.5 | 5 | 15 | 5 |
| Trimethylbenzene 1,2,3- | 526-73-8 | C ₉ H ₁₂ | TrMeBnz | 0.49 | 25 | 75 | 25 |
| Trimethylbenzene,1,2,4- | 95-63-6 | C ₉ H ₁₂ | TrMeBnz | 0.43 | 25 | 75 | 25 |
| Trimethylbenzene,1,3,5- | 108-67-8 | C ₉ H ₁₂ | TrMeBnz | 0.34 | 25 | 75 | 25 |
| Turpentine Crude | | -- | TurptnS | 1 | 20 | 60 | 20 |
| Turpentine Pure | 8006-64-2 | -- | Turptn | 0.45 | 20 | 60 | 20 |
| Vinyl Acetate | 108-05-4 | C ₄ H ₆ O ₂ | VinAcet | 1.1 | 10 | 15 | 10 |
| Vinylidene Chloride | 75-35-4 | C ₂ H ₂ Cl ₂ | VindnCl | 0.95 | 5 | 15 | 5 |
| Xylene, m- | 108-38-3 | C ₈ H ₁₀ | Xylene | 0.44 | 100 | 150 | 100 |
| Xylene, o- | 95-47-6 | C ₈ H ₁₀ | Xylene | 0.6 | 100 | 150 | 100 |
| Xylene, p- | 106-42-3 | C ₈ H ₁₀ | Xylene | 0.46 | 100 | 150 | 100 |

Note – An alarm setpoint of zero corresponds to the alarm being disabled. This is because there is no established OSHA PEL (permissible exposure limit) established for these materials.